Installing OSi Software onto i-STAT Alinity v Using a USB Drive



Click This Link
> Select Download Software

Navigate to Download Your i-STAT Alinity v Handheld Analyzer Software

> Select Downloadable i-STAT Alinity v OSiXX System Software

Save software to a USB drive

** Requires FAT32 formatted USB 2.0 memory stick*

** File must be saved on the top-level directory of the USB drive*

Place instrument into the Base Station and power on the instrument



Insert USB drive into the USB port in the Base Station

From the Home Screen:
> More Options
> Instrument Options
> Software Installation
> Install from USB

Follow prompts until transfer is complete

SUCCESS

OSi Software Installation Results
SUCCESS

FAILURE

OSi Software Installation Results
FAILURE

If additional assistance is required, please contact Zoetis Diagnostic Technical Support at 1-800-822-2947. Support is available 24 hours/day, 7 days/week.

All trademarks are the property of Zoetis Services LLC or a related company or a licensor unless otherwise noted. ©2022 Zoetis Services LLC. All rights reseved. ABX-00247 R1