

i-STAT Crea Cartridge



NAME

i-STAT Crea Cartridge – REF 03P84-25

INTENDED USE

The i-STAT Crea cartridge with the i-STAT 1 System is intended for use in the *in vitro* quantification of creatinine in arterial, venous, or capillary whole blood.

Creatinine measurements are used in the diagnosis and treatment of renal diseases, in monitoring renal dialysis, and as a calculation basis for measuring other urine analytes.

SUMMARY AND EXPLANATION/CLINICAL SIGNIFICANCE

Measured:

Creatinine (Crea)

Elevated levels of creatinine are mainly associated with abnormal renal function and occur whenever there is a significant reduction in glomerular filtration rate or when urine elimination is obstructed. The concentration of creatinine is a better indicator of renal function than urea or uric acid because it is not affected by diet, exercise, or hormones. The creatinine level has been used in combination with BUN to differentiate between prerenal and renal causes of an elevated urea/BUN.


TEST PRINCIPLE

The i-STAT System uses direct (undiluted) electrochemical methods. Values obtained by direct methods may differ from those obtained by indirect (diluted) methods. ¹

Measured:

Creatinine (Crea)

Creatinine is measured amperometrically. It is hydrolyzed to creatine in a reaction catalyzed by the enzyme creatinine amidohydrolase. Creatine is then hydrolyzed to sarcosine by creatine amidohydrolase. The oxidation of sarcosine, catalyzed by sarcosine oxidase, produces hydrogen peroxide (H₂O₂). The liberated hydrogen peroxide is oxidized at the platinum electrode to produce a current which is proportional to the sample creatinine concentration.



See below for information on factors affecting results. Certain substances, such as drugs, may affect analyte levels *in vivo*.² If results appear inconsistent with the clinical assessment, the patient sample should be retested using another cartridge.

REAGENTS

Contents

Each i-STAT cartridge contains one reference electrode (when potentiometric sensors are included in the cartridge configuration), sensors for the measurement of specific analytes, and a buffered aqueous calibrant solution that contains known concentrations of analytes and preservatives. A list of reactive ingredients for the i-STAT Creatinine cartridge is shown below:

Sensor	Reactive Ingredient	Biological Source	Minimum Quantity
Crea	Creatinine	N/A	158.4 µmol/L
	Creatine Amidinohydrolase	<i>Actinobacillus</i> spp.	0.01 IU
	Creatinine Amidohydrolase	Microbial	0.02 IU
	Sarcosine Oxidase	Microbial	0.001 IU

Warnings and Precautions

- For *in vitro* diagnostic use.
- Cartridges are intended for single-use only. Do not reuse.
- Refer to the i-STAT 1 System Manual for all warnings and precautions.

Storage Conditions

- Refrigeration at 2–8 °C (35–46 °F) until expiration date.
- Room Temperature at 18–30 °C (64–86 °F). Refer to the cartridge box for recommended shelf life.

INSTRUMENTS

The i-STAT Creatinine cartridge is intended for use with i-STAT 1 analyzer.

SPECIMEN COLLECTION AND PREPARATION FOR ANALYSIS

Specimen Types

Arterial, venous or capillary whole blood.

Sample volume: 65 µL

Blood Collection Options and Test Timing (time from collection to cartridge fill)

Analyte	Syringes	Test Timing	Evacuated Tubes	Test Timing	Capillary Tubes	Test Timing
Creatinine	Without anticoagulant	3 minutes	Without anticoagulant	3 minutes	With balanced heparin anticoagulant or lithium heparin if labeled for the measurement of electrolytes	3 minutes
	With balanced heparin anticoagulant or lithium heparin anticoagulant (syringe must be filled per manufacturer's recommendation)	30 minutes	With lithium heparin anticoagulant (tubes must be filled per manufacturer's recommendation) <ul style="list-style-type: none"> • Remix thoroughly 	30 minutes		

Analyte	Syringes	Test Timing	Evacuated Tubes	Test Timing	Capillary Tubes	Test Timing
	<ul style="list-style-type: none"> Remix thoroughly before filling cartridge. 		before filling cartridge.			

PROCEDURE FOR CARTRIDGE TESTING

Each cartridge is sealed in a foil pouch for protection during storage--do not use if pouch has been punctured.

- A cartridge should not be removed from its protective pouch until it is at room temperature (18-30 °C or 64-86 °F). For best results, the cartridge and analyzer should be at room temperature.
- Since condensation on a cold cartridge may prevent proper contact with the analyzer, allow refrigerated cartridges to equilibrate at room temperature for 5 minutes for a single cartridge and 1 hour for an entire box before use.
- Use a cartridge immediately after removing it from its protective pouch. Prolonged exposure may cause a cartridge to fail a Quality Check.
- Do not return unopened, previously refrigerated cartridges to the refrigerator.
- Cartridges may be stored at room temperature for the time frame indicated on the cartridge box.

Filling and Sealing the Cartridge (after cartridge has been equilibrated and blood sample has been collected)

- Place the cartridge on a flat surface.
- Mix the sample thoroughly. Invert a lithium heparin blood collection tube at least 10 times. If sample was collected into a syringe, invert syringe for 5 seconds then roll the syringe between the palms (hands parallel to the ground) for 5 seconds, flip and roll for an additional 5 seconds. The blood in the hub of the syringe will not mix, therefore expelling 2 drops before filling a cartridge is desired. Note that it may be difficult to properly mix a sample in a 1.0 mL syringe.
- Fill the cartridge immediately after mixing. Direct the hub of syringe or tip of the transfer device (capillary tube, pipette, or dispensing tip) into the sample well of the cartridge.
- Slowly dispense sample into the sample well until the sample reaches the fill mark indicated on the cartridge. Cartridge is properly filled when the sample reaches the 'fill to' mark and a small amount of sample is in the sample well. The sample should be continuous, no bubbles or breaks (see System Manual for details).
- Fold the snap closure over the sample well.

Performing Patient Analysis

- Press the power button to turn on the handheld.
- Press 2 for *i-STAT Cartridge*.
- Follow the handheld prompts.
- Scan the lot number on the cartridge pouch.
- Continue normal procedures for preparing the sample, and filling and sealing the cartridge.
- Push the sealed cartridge into the handheld port until it clicks into place. Wait for the test to complete.
- Review the results.

For additional information for cartridge testing, refer to the i-STAT 1 System Manual located at www.pointofcare.abbott.

Analysis Time

Approximately 130–200 seconds.

Quality Control

The i-STAT quality control regimen comprises four aspects, with a system design that reduces the opportunity for error, including:

1. A series of automated, on-line quality measurements that monitors the sensors, fluidics, and instrumentation each time a test is performed.
2. A series of automated, on-line procedural checks that monitors the user each time a test is performed.
3. Liquid materials are available to be used to verify the performance of a batch of cartridges when they are first received or when storage conditions are in question. The performance of this procedure is not a manufacturer's system instruction.
4. Traditional quality control measurements that verify the instrumentation using an independent device, which simulates the characteristics of the electrochemical sensors in a way that stresses the performance characteristics of the instrumentation.

For additional information on Quality Control, refer to the i-STAT 1 System Manual located at www.pointofcare.abbott.

Calibration Verification

Calibration Verification is a procedure intended to verify the accuracy of results over the entire measurement range of a test. The performance of this procedure is not a manufacturer's system instruction. However, it may be required by regulatory or accreditation bodies. While the Calibration Verification Set contains five levels, verification of the measurement range could be accomplished using the lowest, highest and mid-levels.

EXPECTED VALUES

TEST	UNITS *	REPORTABLE RANGE	REFERENCE RANGE	
			<i>arterial</i>	<i>venous</i>
MEASURED				
Crea	mg/dL	0.2–20.0	0.6–1.3 ³	
	µmol/L	18–1768	53–115	

* The i-STAT System can be configured with the preferred units. (See "Unit Conversion" below.)

Unit Conversion

- o **Creatinine (Crea):** To convert mg/dL to µmol/L, multiply the mg/dL value by 88.4.

The reference ranges programmed into the analyzer and shown above are intended to be used as guides for the interpretation of results. Since reference ranges may vary with demographic factors such as age, gender and heritage, it is recommended that reference ranges be determined for the population being tested.

METROLOGICAL TRACEABILITY

The measured analyte in the i-STAT Crea cartridge is traceable to the following reference materials or methods. The i-STAT System controls and calibration verification materials are validated for use only with the i-STAT System and assigned values may not be commutable with other methods.

Creatinine (Crea)

The i-STAT System test for creatinine measures creatinine amount-of-substance concentration in the

plasma fraction of arterial, venous, or capillary whole blood (dimension $\mu\text{mol L}^{-1}$) for in vitro diagnostic use. Creatinine values assigned to i-STAT System controls and calibration verification materials are traceable to the U.S. National Institute of Standards and Technology (NIST) standard reference material SRM967.

Additional information regarding metrological traceability is available from Abbott Point of Care Inc.

PERFORMANCE CHARACTERISTICS

The typical performance data summarized below were collected in health care facilities by health care professionals trained in the use of the i-STAT System and comparative methods. Clinical settings vary and some may require different performance characteristics to assess renal function status than others (e.g., medication dosing, intravenous contrast use, and outpatient clinic). If deemed necessary by a health care facility, performance data should be obtained in specific clinical settings to assure patients' needs are met.

Precision

Precision data was collected in multiple sites as follows: Duplicates of each control fluid were tested in the morning and in the afternoon on five days for a total of 20 replicates. The averaged statistics are presented below.

Test	Units	Aqueous Control	Mean	SD (Standard Deviation)	CV (%) [Coefficient of Variation (%)]
Crea	mg/dL	Level 1	4.33	0.131	3.0
		Level 3	0.81	0.039	4.8

Method Comparison

Method comparison data were collected using CLSI guideline EP9-A. ⁴

Deming regression analysis ⁵ was performed on the first replicate of each sample set. In the method comparison table, n is the number of specimens in the data set, Sxx and Syy refer to estimates of imprecision based on the duplicates of the comparative and the i-STAT methods respectively, Sy.x is the standard error of the estimate, and r is the correlation coefficient. *

Method comparisons will vary from site to site due to differences in sample handling, comparative method calibration and other site-specific variables.

* The usual warning relating to the use of regression analysis is summarized here as a reminder. For any analyte, "if the data is collected over a narrow range, the estimate of the regression parameters are relatively imprecise and may be biased. Therefore, predictions made from these estimates may be invalid". ⁵ The correlation coefficient, r, can be used as a guide to assess the adequacy of the comparative method range in overcoming this problem, and, as a guide, the range of data can be considered adequate for $r > 0.975$.

Creatinine/Crea (mg/dL)		Roche Integra 800	Beckman LX20®	J & J Vitros 950	Dade Dimension RxL
Venous blood samples, collected in lithium or sodium heparin Vacutainer® tubes, and arterial blood samples, collected in blood gas syringes, were analyzed in duplicate on the i-STAT System. A portion of each specimen was centrifuged and the separated plasma was analyzed on the comparative method.	n	30	58	31	36
	Sxx	0.029	0.141	0.04	0.04
	Syy	0.112	0.143	0.12	0.06
	Slope	0.929	0.960	0.948	0.964
	Int't	0.237	0.022	0.206	0.100
	Sy.x	0.204	0.261	0.165	0.123
	Xmin	0.4	0.7	0.5	0.5
	Xmax	10.3	20.0	7.2	5.7
	r	0.997	0.996	0.991	0.986

FACTORS AFFECTING RESULTS

The following substances were evaluated in plasma for relevant analyte at the test concentrations recommended in CLSI guideline EP7-A2⁶ unless otherwise noted. For those identified as an interferant the interference is described.

Substance	Test Concentration (mmol/L)	Analyte	Interference (Yes/No)	Comment
Acetaldehyde	0.04 ⁷	Crea	No	
Acetaminophen	1.32	Crea	Yes	Increased results
Acetaminophen (therapeutic)	0.132 ⁷	Crea	No	
Acetylcysteine	10.2	Crea	Yes	Increased results
Acetylcysteine (therapeutic)	0.3 ^{8,9}	Crea	No	
Ascorbate	0.34	Crea	Yes	Increased by up to 0.3 mg/dL
Bicarbonate	35.0	Crea	No	
Bilirubin	0.342	Crea	No	
Bromide (therapeutic)	2.5 ^{10,11,12}	Crea	Yes	Increased results
Calcium Chloride	5.0	Crea	No	
Creatine	0.382	Crea	Yes	Increased by up to 0.3 mg/dL. See Other Factors Affecting Results below for CO ₂ dependence
Dopamine	0.006	Crea	No	
Formaldehyde	0.133 ⁷	Crea	No	
β-Hydroxybutyrate	6.0 ¹³	Crea	No	
Glycolic Acid	10.0	Crea	Yes	Decreased results. Use another method.
Hydroxyurea	0.92	Crea	Yes	Increased results. Use another method.
Lactate	6.6	Crea	No	
Methyldopa	0.071	Crea	No	
Nithiodote (Sodium thiosulfate)	16.7 ¹⁴	Crea	Yes	Increased results
Pyruvate	0.31	Crea	No	
Salicylate	4.34	Crea	No	
Uric Acid	1.4	Crea	No	

The degree of interference at concentrations other than those reported above might not be predictable. It is possible that interfering substances other than those tested may be encountered.

Relevant comments regarding interference of Acetaminophen, Acetylcysteine, Bromide, Hydroxyurea and Nithiodote are noted below:

- Acetaminophen has been shown to interfere with i-STAT creatinine results at a 1.32 mmol/L, a toxic concentration that is proscribed by the CLSI guideline. Acetaminophen at 0.132 mmol/L, which represents the upper end of the therapeutic concentration range, has been shown not to significantly interfere with i-STAT creatinine results.
- Acetylcysteine has been tested at two levels: the CLSI recommended level of 10.2 mmol/L and a concentration of 0.30 mmol/L. The latter is 3 times the peak plasma therapeutic concentration associated with treatment to reverse acetaminophen poisoning. APOC has not identified a therapeutic condition that would lead to levels consistent with the CLSI recommended level.

Acetylcysteine at a concentration of 10.2 mmol/L increased i-STAT creatinine results, while acetylcysteine at a concentration of 0.3 mmol/L did not significantly interfere with i-STAT creatinine results.








- Bromide has been tested at two levels: the CLSI recommended level and a therapeutic plasma concentration level of 2.5 mmol/L. The latter is the peak plasma concentration associated with halothane anesthesia, in which bromide is released. APOC has not identified a therapeutic condition that would lead to levels consistent with the CLSI recommended level. Bromide tested at concentrations of 2.5 and 37.5 mmol/L interfered with i-STAT creatinine results.
- Hydroxyurea is a DNA synthesis inhibitor used in the treatment of various forms of cancer, sickle cell anemia, and HIV infection. This drug is used to treat malignancies including melanoma, metastatic ovarian cancer, and chronic myelogenous leukemia. It is also used in the treatment of polycythemia vera, thrombocytopenia, and psoriasis. At typical doses ranging from 500 mg to 2 g/day, concentrations of hydroxyurea in patients' blood may be sustained at approximately 100 to 500 µmol/L. Higher concentrations may be observed soon after dosing or at higher therapeutic doses.
- Nithiodote (sodium thiosulfate) is indicated for the treatment of acute cyanide poisoning. The journal article titled "Falsely increased chloride and missed anion gap elevation during treatment with sodium thiosulfate" indicated that sodium thiosulfate could be used in the treatment of calciphylaxis indicating that "the highest concentration likely to be seen in plasma [is] after infusion of a 12.5 g dose of sodium thiosulfate pentahydrate. Assuming that the 12.5 g dose of sodium thiosulfate pentahydrate is distributed in a typical blood volume of 5 L with a hematocrit of 40%, the peak sodium thiosulfate plasma concentration expected is 16.7 mmol/L."¹⁴

*It is possible that other interfering substance may be encountered. The degree of interference at concentrations other than those listed might not be predictable.

OTHER FACTORS AFFECTING RESULTS

Factor	Analyte	Effect
Creatine	Creatinine	The normal range of creatine concentration in plasma is 0.17–0.70 mg/dL (13 – 53 µmol/L) in males and 0.35 – 0.93 mg/dL (27 – 71 µmol/L) in females. ⁷ Creatine may be elevated in patients using creatine supplements, experiencing muscle trauma or other primary or secondary myopathies, taking statins for hyperlipidemia control, or in patients with hyperthyroidism or a rare genetic defect of the creatine transporter protein.
CO ₂ dependence	Creatinine	The dependence of the i-STAT creatinine with respect to Carbon Dioxide (CO ₂) is as follows: For creatinine results ≤ 2.0 mg/dL, no correction for <i>PCO</i> ₂ is required. For creatinine results > 2.0 mg/dL, the following correction applies: Creatinine_{corrected} = creatinine * (1 + 0.0025 * (PCO₂ – 40))

KEY TO SYMBOLS

Symbol	Definition/Use
14 	14 days room temperature storage at 18–30 °C.
	Use by or expiration date. The expiration date, expressed as YYYY-MM-DD, indicates the last day the product may be used.
LOT	Manufacturer's lot number or batch code. The lot number or batch code appears adjacent to this symbol.
	Sufficient for <n> tests.
EC REP	Authorized representative for Regulatory Affairs in the European Community.
	Temperature limitations. The upper and lower limits for storage are adjacent to upper and lower arms.
REF	Catalog number, list number, or reference.
	Do not reuse.
	Manufacturer.
	Consult instructions for use or see System Manual for instructions.
IVD	<i>In vitro</i> diagnostic medical device.
CE	Compliance to the European directive on <i>in vitro</i> diagnostic devices (98/79/EC)
Rx ONLY	For prescription use only.

Additional Information: To obtain additional product information and technical support, refer to the Abbott company website at www.pointofcare.abbott.

References

1. Tietz NW, Pruden EL, Siggaard-Andersen O. Electrolytes. In: C.A. Burtis and E.R. Ashwood, ed. *Tietz Textbook of Clinical Chemistry*. Second Edition ed. Philadelphia: W.B. Saunders Company; 1994.
2. Young DS. *Effects of Drugs on Clinical Laboratory Tests*. 3rd ed. ed. Washington, DC: American Association of Clinical Chemistry; 1990.
3. Burtis CA, Ashwood ER, Bruns DE, eds. *Tietz Textbook of Clinical Chemistry and Molecular Diagnostics*. 4th Edition ed: Elsevier Saunders Inc.; 2006.
4. CLSI. Method Comparison and Bias Estimation Using Patient Samples; Approved Guideline. *CLSI document EP9-A*. 1995.
5. Cornbleet PJ, Gochman N. Incorrect least-squares regression coefficients in method-comparison analysis. *Clinical Chemistry*. 1979;25(3).
6. Clinical and Laboratory Standards Institute. Interference Testing in Clinical Chemistry; Approved Guideline—Second Edition. *CLSI document EP7-A2*. 2005.
7. Wu AHB. *Tietz Clinical Guide to Laboratory Tests*: Elsevier Health Sciences; 2006.
8. Whillier S, Raftos JE, Chapman B, Kuchel PW. Role of N-acetylcysteine and cystine in glutathione synthesis in human erythrocytes. *Redox Report*. 2009;14(3):115-121.
9. Ventura P, Panini R, Pasini MC, Scarpetta G, Salvioli G. N-acetyl-cysteine reduces homocysteine plasma levels after single intravenous administration by increasing thiols urinary excretion. *Pharmacological Research*. 1999;40(4):345-350.
10. Hankins DC, Kharasch ED. Determination of the halothane metabolites trifluoroacetic acid and bromide in plasma and urine by ion chromatography. *Journal of Chromatography B: Biomedical Applications*. May 1997;692(2):413-418.
11. Kharasch ED, Hankins D, Mautz D, Thummel KE. Identification of the enzyme responsible for oxidative halothane metabolism: Implications for prevention of halothane hepatitis. *Lancet*. May 1996;347(9012):1367-1371.
12. Morrison JE, Friesen RH. Elevated serum bromide concentrations following repeated halothane anaesthesia in a child. *Canadian Journal of Anaesthesia*. October 1990;37(7):801-803.
13. Charles RA, Bee YM, Eng PHK, Goh SY. Point-of-care blood ketone testing: Screening for diabetic ketoacidosis at the emergency department. *Singapore Medical Journal*. November 2007;48(11):986-989.
14. Wendroth SM, Heady TN, Haverstick DM, et al. Falsely increased chloride and missed anion gap elevation during treatment with sodium thiosulfate. *Clinica Chimica Acta*. April 2014;431:77-79.

i-STAT is a trademark of the Abbott group of companies.

Vacutainer is a trademark of Becton Dickinson and Company, Franklin Lakes, NJ USA.

CX®3, LX20, CX9, Coulter S Plus are trademark of Beckman Coulter Incorporated, Fullerton, CA USA.

Ektachem was a trademark of Kodak Clinical Diagnostics. This system is now the Vitros® distributed by Ortho-Clinical Diagnostics, Rochester, NY, USA.

Stat Profile is a trademark of Nova Biomedical, Waltham, MA USA.


ICA 1 is a trademark of Radiometer Medical A/S, Copenhagen, Denmark.

The Bayer 860 analyzer is manufactured by Bayer Diagnostics, Tarrytown, NY USA.

Dimension RxL-Xpand is a trademark of Dade Behring Inc., Deerfield, IL USA.

Cell-Dyn is a trademark of Abbott Laboratories, Abbott Park, IL USA.

SE9500 is a trademark of Sysmex America Inc., Mundelein, IL USA.



©2019 Abbott Point of Care Inc. All rights reserved. Printed in USA.